



校准证书

Calibration Certificate

证书(记录)编号: 085006180
Certificate No

第 3 页, 共 3 页
Page 3 of 3 Pages

校准结果

Results of Calibration

- 外观及功能检查: 合格。
- 示值误差如下表:

通道号	被校温度点 (°C)							
	0	50	100	150	200	250	300	350
	指示基本误差 (°C)							
01	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
02	0.0	+0.2	+0.1	+0.1	+0.1	+0.1	+0.1	+0.1
03	0.0	+0.2	+0.1	+0.1	0.0	0.0	-0.2	-0.3
04	0.0	0.0	-0.1	0.0	0.0	0.0	+0.1	+0.1
05	0.0	+0.3	+0.2	+0.2	+0.1	0.0	-0.1	-0.1
06	0.0	+0.3	+0.2	+0.2	+0.1	+0.1	0.0	-0.1

附注:

- 本次校准结果不确定度(扩展不确定度) $l=0.4^{\circ}\text{C}$, $k=2$ 。
- 说明书中技术要求, 允许误差: $\pm 2^{\circ}\text{C}$ 。

(以下空白)



校准员: 喻晓虎
Calibrated by

核验员: 孙世海
Checked by